

Control Number _____
Serial Number _____
Date _____

Ellman Surgitron

Settings	P - P Voltage Expected	P - P Actual	Watts Expected	Watts Actual
5	245		15	
10	323		26	
15	380		36	
20	424		45	
25	461		53	
30	490		60	
40	540		73	
50	583		85	
60	617		95	
70	642		103	
80	660		109	
90	678		115	
100	693		120	